

**RFP 2019 1219 001**  
**Cardiac Monitor Defibrillator**  
**Bid Tabulation**  
**December 19, 2019**

**#1 Stryker – signed by the bidder**

**Total Bid Amount \$85, 838**

**Trade-In per unit \$4,000 (Calculated in the Total Bid Amount)**

**Exceptions:** No real time 12 Lead ECG has “Continuous ST segment tracking” with auto print, No real time CPR feedback / No external battery charger

**Total Cost per Machine: \$28,612.67**

**#2 ZOLL – signed by the bidder**

**Total Bid Amount \$89,800.98**

**Trade-In per unit \$4,200 (Calculated in the Total Bid Amount)**

**Exceptions:** None

**Total Cost per Machine: \$29,933.66**

Prepared by :

Joe King  
Joe King, Morgan County Ambulance Service Director

Awarded:

ZOLL

**MORGAN COUNTY, COLORADO**

Chairman

James Zumbly

Date

1-7-2020

Commissioner

Mark R. Hark

Date


1-7-2020

  
\_\_\_\_\_  
Commissioner

1.7.2020  
\_\_\_\_\_  
Date



**ATTEST:**

  
\_\_\_\_\_  
Susan L. Bailey, Clerk to the Board *the Board*

(seal)  
Date: 1.7.2020  
\_\_\_\_\_